



Pace Analytical Services, LLC  
575 Broad Hollow Road  
Melville, NY 11747  
(631)694-3040

June 27, 2018

Mark Ruhnke  
Eisenbach & Ruhnke  
291 Genesee Street  
Utica, NY 13501

RE: Project: GREENBURGH NORTH CASTLE  
Pace Project No.: 7056064

Dear Mark Ruhnke:

Enclosed are the analytical results for sample(s) received by the laboratory on June 23, 2018. The results relate only to the samples included in this report. Results reported herein conform to the most current, applicable TNI/NELAC standards and the laboratory's Quality Assurance Manual, where applicable, unless otherwise noted in the body of the report.

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

James Murphy  
james.murphy@pacelabs.com  
(518)346-4592  
Project Manager

Enclosures

cc: Larry Kinne, Eisenbach & Ruhnke



### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



Pace Analytical Services, LLC  
575 Broad Hollow Road  
Melville, NY 11747  
(631)694-3040

## CERTIFICATIONS

Project: GREENBURGH NORTH CASTLE  
Pace Project No.: 7056064

---

### Long Island Certification IDs

575 Broad Hollow Rd, Melville, NY 11747  
New York Certification #: 10478 Primary Accrediting Body  
New Jersey Certification #: NY158  
Pennsylvania Certification #: 68-00350  
Connecticut Certification #: PH-0435

Maryland Certification #: 208  
Rhode Island Certification #: LAO00340  
Massachusetts Certification #: M-NY026  
New Hampshire Certification #: 2987

---

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

### SAMPLE SUMMARY

Project: GREENBURGH NORTH CASTLE  
Pace Project No.: 7056064

Lab ID	Sample ID	Matrix	Date Collected	Date Received
7056064001	R1-BOYS BATHROOM SINK 1 (LEFT)	Drinking Water	06/22/18 00:00	06/23/18 10:30
7056064002	R2-BOYS BATHROOM SINK 2 (RIGHT)	Drinking Water	06/22/18 00:00	06/23/18 10:30
7056064003	R3-GIRLS BATHROOM SINK 1 (LEFT)	Drinking Water	06/22/18 00:00	06/23/18 10:30
7056064004	R4-GIRLS BATHROOM SINK 2 (RIGHT)	Drinking Water	06/22/18 00:00	06/23/18 10:30
7056064005	R5-KITCHEN SINK	Drinking Water	06/22/18 00:00	06/23/18 10:30
7056064006	R8-FOUNTAIN MAIN FLOOR	Drinking Water	06/22/18 00:00	06/23/18 10:30
7056064007	R10-BATHROOM SINK-LOWER LEVEL	Drinking Water	06/22/18 00:00	06/23/18 10:30
7056064008	R11-PRINCIPAL OFFICE-SINK	Drinking Water	06/22/18 00:00	06/23/18 10:30

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



### SAMPLE ANALYTE COUNT

Project: GREENBURGH NORTH CASTLE  
Pace Project No.: 7056064

---

Lab ID	Sample ID	Method	Analysts	Analytes Reported
7056064001	R1-BOYS BATHROOM SINK 1 (LEFT)	EPA 200.8	SK2	1
7056064002	R2-BOYS BATHROOM SINK 2 (RIGHT)	EPA 200.8	SK2	1
7056064003	R3-GIRLS BATHROOM SINK 1 LEFT	EPA 200.8	SK2	1
7056064004	R4-GIRLS BATHROOM SINK 2 RIGHT	EPA 200.8	SK2	1
7056064005	R5-KITCHEN SINK	EPA 200.8	SK2	1
7056064006	R8-FOUNTAIN MAIN FLOOR	EPA 200.8	SK2	1
7056064007	R10-BATHROOM SINK-LOWER LEVEL	EPA 200.8	SK2	1
7056064008	R11-PRINCIPAL OFFICE-SINK	EPA 200.8	SK2	1

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



### ANALYTICAL RESULTS

Project: GREENBURGH NORTH CASTLE  
Pace Project No.: 7056064

Sample: R1-BOYS BATHROOM SINK 1 (LEFT)	Lab ID: 7056064001	Collected: 06/22/18 00:00	Received: 06/23/18 10:30	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>	Analytical Method: EPA 200.8							
Lead	1.3	ug/L	1.0	1		06/26/18 17:23	7439-92-1	

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



### ANALYTICAL RESULTS

Project: GREENBURGH NORTH CASTLE  
Pace Project No.: 7056064

Sample:	Lab ID:	Collected:	Received:	Matrix:				
R2-BOYS BATHROOM SINK 2 (RIGHT)	7056064002	06/22/18 00:00	06/23/18 10:30	Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8							
Lead	<1.0	ug/L	1.0	1		06/26/18 17:26	7439-92-1	

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



### ANALYTICAL RESULTS

Project: GREENBURGH NORTH CASTLE  
Pace Project No.: 7056064

Sample:	Lab ID:	Collected:	Received:	Matrix:				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>R3-GIRLS BATHROOM SINK 1 LEFT</b>	<b>7056064003</b>	06/22/18 00:00	06/23/18 10:30	Drinking Water				
<b>200.8 MET ICPMS Drinking Water</b>	Analytical Method: EPA 200.8							
Lead	<1.0	ug/L	1.0	1		06/26/18 17:35	7439-92-1	

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



Pace Analytical Services, LLC  
575 Broad Hollow Road  
Melville, NY 11747  
(631)694-3040

### ANALYTICAL RESULTS

Project: GREENBURGH NORTH CASTLE

Pace Project No.: 7056064

Sample: R4-GIRLS BATHROOM SINK 2 RIGHT    Lab ID: 7056064004    Collected: 06/22/18 00:00    Received: 06/23/18 10:30    Matrix: Drinking Water

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>	Analytical Method: EPA 200.8							
Lead	1.4	ug/L	1.0	1		06/26/18 17:38	7439-92-1	

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.





### ANALYTICAL RESULTS

Project: GREENBURGH NORTH CASTLE  
Pace Project No.: 7056064

Sample: R5-KITCHEN SINK	Lab ID: 7056064005	Collected: 06/22/18 00:00	Received: 06/23/18 10:30	Matrix: Drinking Water					
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8								
Lead	<1.0	ug/L	1.0	1		06/26/18 17:41	7439-92-1		

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



### ANALYTICAL RESULTS

Project: GREENBURGH NORTH CASTLE  
Pace Project No.: 7056064

Sample: R8-FOUNTAIN MAIN FLOOR	Lab ID: 7056064006	Collected: 06/22/18 00:00	Received: 06/23/18 10:30	Matrix: Drinking Water					
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
<b>200.8 MET ICPMS Drinking Water</b>	Analytical Method: EPA 200.8								
Lead	<1.0	ug/L	1.0	1		06/26/18 17:44	7439-92-1		

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



Pace Analytical Services, LLC  
575 Broad Hollow Road  
Melville, NY 11747  
(631)694-3040

### ANALYTICAL RESULTS

Project: GREENBURGH NORTH CASTLE  
Pace Project No.: 7056064

Sample: R10-BATHROOM SINK-LOWER LEVEL      Lab ID: 7056064007      Collected: 06/22/18 00:00      Received: 06/23/18 10:30      Matrix: Drinking Water

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>	Analytical Method: EPA 200.8							
Lead	<1.0	ug/L	1.0	1		06/26/18 17:47	7439-92-1	

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



Pace Analytical Services, LLC  
575 Broad Hollow Road  
Melville, NY 11747  
(631)694-3040

### ANALYTICAL RESULTS

Project: GREENBURGH NORTH CASTLE  
Pace Project No.: 7056064

Sample: R11-PRINCIPAL OFFICE-SINK    Lab ID: 7056064008    Collected: 06/22/18 00:00    Received: 06/23/18 10:30    Matrix: Drinking Water

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>200.8 MET ICPMS Drinking Water</b>		Analytical Method: EPA 200.8						
Lead	2.3	ug/L	1.0	1		06/26/18 17:50	7439-92-1	

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

**QUALITY CONTROL DATA**

Project: GREENBURGH NORTH CASTLE  
Pace Project No.: 7056064

QC Batch: 72970 Analysis Method: EPA 200.8  
QC Batch Method: EPA 200.8 Analysis Description: 200.8 MET No Prep Drinking Water  
Associated Lab Samples: 7056064001, 7056064002, 7056064003, 7056064004, 7056064005, 7056064006, 7056064007, 7056064008

METHOD BLANK: 334786 Matrix: Water  
Associated Lab Samples: 7056064001, 7056064002, 7056064003, 7056064004, 7056064005, 7056064006, 7056064007, 7056064008

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	06/26/18 16:31	

LABORATORY CONTROL SAMPLE: 334787

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.2	100	85-115	

MATRIX SPIKE SAMPLE: 334789

Parameter	Units	7056063001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	2.5	2	4.7	106	70-130	

MATRIX SPIKE SAMPLE: 334791

Parameter	Units	7056063002 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	2.7	2	5.0	115	70-130	

SAMPLE DUPLICATE: 334788

Parameter	Units	7056063001 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	2.5	2.4	6	20	

SAMPLE DUPLICATE: 334790

Parameter	Units	7056063002 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	2.7	2.6	2	20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

**REPORT OF LABORATORY ANALYSIS**

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.

## QUALIFIERS

Project: GREENBURGH NORTH CASTLE  
Pace Project No.: 7056064

---

### DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.  
ND - Not Detected at or above adjusted reporting limit.  
TNTC - Too Numerous To Count  
J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.  
MDL - Adjusted Method Detection Limit.  
PQL - Practical Quantitation Limit.  
RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.  
S - Surrogate  
1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.  
Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.  
LCS(D) - Laboratory Control Sample (Duplicate)  
MS(D) - Matrix Spike (Duplicate)  
DUP - Sample Duplicate  
RPD - Relative Percent Difference  
NC - Not Calculable.  
SG - Silica Gel - Clean-Up  
U - Indicates the compound was analyzed for, but not detected.  
N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.  
Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.  
TNI - The NELAC Institute.

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

**QUALITY CONTROL DATA CROSS REFERENCE TABLE**

Project: GREENBURGH NORTH CASTLE  
Pace Project No.: 7056064

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
7056064001	R1-BOYS BATHROOM SINK 1 (LEFT)	EPA 200.8	72970		
7056064002	R2-BOYS BATHROOM SINK 2 (RIGHT)	EPA 200.8	72970		
7056064003	R3-GIRLS BATHROOM SINK 1 LEFT	EPA 200.8	72970		
7056064004	R4-GIRLS BATHROOM SINK 2 RIGHT	EPA 200.8	72970		
7056064005	R5-KITCHEN SINK	EPA 200.8	72970		
7056064006	R8-FOUNTAIN MAIN FLOOR	EPA 200.8	72970		
7056064007	R10-BATHROOM SINK-LOWER LEVEL	EPA 200.8	72970		
7056064008	R11-PRINCIPAL OFFICE-SINK	EPA 200.8	72970		

**REPORT OF LABORATORY ANALYSIS**

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.

WO#: 7056064



7056064

Utica

315-735-1916

CHAIN OF CUSTODY

Invoice to:

Eisenbach and Ruhne Engineering, P.C.

Attn: *Joelle Smith - JSmith@erengpc.com*

Project # 35-18-01

PO #

Contact Person: *Mark Ruhnke*

email Results to: *MRuhnke@erengpc.com*

Project State NY

Analysis Due Date:

TAT:

PROJECT NAME: *Greenburgh North Castle - Reach Academy*

ANALYSIS

LAB ID - FOR LAB USE ONLY	CLIENT SAMPLE ID	SOIL (S/L)/WATER (W)	DATE	Time	Analysis	
					Analysis	Analysis
	R1 - Boys Bathroom Sink 1 (left)	W	6/22/18		X	
	R2 - Boys Bathroom Sink 2 (right)	W			X	
	R3 - Girls Bathroom Sink 1 (left)	W			X	
	R4 - Girls Bathroom Sink 2 (right)	W			X	
	R5 - Kitchen Sink	W			X	
	R8 - Fountain Main Floor	W			X	
	R10 - Bathroom Sink - Lower Level	W			X	
	R11 - Principal Office - Sink	W			X	

EPA Pb 200.8

NOTES:

SAMPLER NAME AND SIGNATURE

PRINT Name of SAMPLER: *Matt Immen*

SIGNATURE of SAMPLER: *[Signature]*

DATE Sampled (MM/DD/YY): 6/22/18

Relinquished BY: *[Signature]*

Accepted BY: *A. DeB...* (PACE)

Relinquished BY: *A. DeB...* (PACE) via FedEx

Accepted BY: *[Signature]*

Date

6/22/18

6/22/18

6/22/18

6/22/18

Time

13:12

13:12

16:00

10:50

Temp in °C

26.2 (18)

Custody Sealed (Y/N)

N

Samples Intact (Y/N)

Y





Sample Condition Upon Receipt

Client Name: Eisenbach + Ruhn

WO#: 7056064

PM: JM2 Due Date: 06/28/18

CLIENT: ER

Courier: [X] Fed Ex [ ] UPS [ ] USPS [ ] Client [ ] Commercial [ ] Pace [ ] Other

Tracking #: 4144 8129 3708

Custody Seal on Cooler/Box Present: [X] Yes [ ] No Seals intact: [X] Yes [ ] No

Packing Material: [ ] Bubble Wrap [ ] Bubble Bags [X] Ziploc [ ] None [ ] Other

Thermometer Used: TH091

Correction Factor: 0.0

Cooler Temperature (°C):

Cooler Temperature Corrected (°C):

Temp should be above freezing to 6.0°C

USDA Regulated Soil [X] N/A, water sample

Date and Initials of person examining contents: JK 6/23/18

Did samples originate from a foreign source (internationally, including Hawaii and Puerto Rico)? [ ] Yes [X] No

Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, IA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)? [ ] YES [X] NO

If Yes to either question, fill out a Regulated Soil Checklist (F-LI-C-010) and include with SCUR/COC paperwork.

				COMMENTS:
Chain of Custody Present:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No		1.
Chain of Custody Filled Out:	<input type="checkbox"/> Yes	<input type="checkbox"/> No		2.
Chain of Custody Relinquished:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No		3.
Sampler Name & Signature on COC:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	<input type="checkbox"/> N/A	4.
Samples Arrived within Hold Time:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No		5.
Short Hold Time Analysis (<72hr):	<input type="checkbox"/> Yes	<input checked="" type="checkbox"/> No		6.
Rush Turn Around Time Requested:	<input type="checkbox"/> Yes	<input checked="" type="checkbox"/> No		7.
Sufficient Volume: (Triple volume provided for MS/MSD):	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No		8.
Correct Containers Used:	<input type="checkbox"/> Yes	<input type="checkbox"/> No		9.
-Pace Containers Used:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No		
Containers Intact:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No		10.
Filtered volume received for Dissolved tests	<input type="checkbox"/> Yes	<input type="checkbox"/> No	<input checked="" type="checkbox"/> N/A	11. Note if sediment is visible in the dissolved container.
Sample Labels match COC:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No		12.
-Includes date/time/ID/Analysis Matrix SL WT OIL				
All containers needing preservation have been checked	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	<input type="checkbox"/> N/A	13. <input type="checkbox"/> HNO <sub>3</sub> <input type="checkbox"/> H <sub>2</sub> SO <sub>4</sub> <input type="checkbox"/> NaOH <input type="checkbox"/> HCl
pH paper Lot # H C 739245				Sample #
All containers needing preservation are found to be in compliance with EPA recommendation? (HNO <sub>3</sub> , H <sub>2</sub> SO <sub>4</sub> , HCl, NaOH > 9 Sulfide, NaOH > 12 Cyanide)	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	<input type="checkbox"/> N/A	Initial when completed: Lot # of added preservative: Date/Time preservative added:
Exceptions: VOA, Coliform, TOC/DOC, Oil and Grease, DRO/BO15 (water). Per Method, VOA pH is checked after analysis				
Samples checked for dechlorination:	<input type="checkbox"/> Yes	<input type="checkbox"/> No	<input checked="" type="checkbox"/> N/A	14. Positive for Res. Chlorine? Y N
KI starch test strips Lot #				
Residual chlorine strips Lot #				
Headspace in VOA Vials (>6mm):	<input type="checkbox"/> Yes	<input type="checkbox"/> No	<input checked="" type="checkbox"/> N/A	15.
Trip Blank Present:	<input type="checkbox"/> Yes	<input type="checkbox"/> No	<input checked="" type="checkbox"/> N/A	16.
Trip Blank Custody Seals Present	<input type="checkbox"/> Yes	<input type="checkbox"/> No	<input checked="" type="checkbox"/> N/A	
Pace Trip Blank Lot # (if applicable):				

Field Data Required? Y / N

Date/Time:

Client Notification/ Resolution:

Person Contacted:

Comments/ Resolution:

